

Tester optic in cutie de plastic - 1980



Descriere

- Utilizare
- \bullet for measuring the concentration of electrolytes in the battery in the range from 1,10 to 1,40 Kg/L
- for measuring the usefulness of coolant (ethylene or propylene) in the range till -50°C
- \bullet for measuring the usefulness of windscreen fluid in the range till -40°C
- scale grating 0,01 Kg / L in measuring the electrolyte concentration and 5 $^{\circ}$ C in the measurement of the other two fluids
- before and after the use the tester needs to be cleaned
- $\ensuremath{\bullet}$ packed in a plastic box and SOS tool tray, which allows safe storage of tester
- supplied with pipette and screwdriver to adjust the calibration
- included instructions for use

Date Tehnice

	ů	1	200	
621403	315	-	-	0

Date de ambalare

	Mod de reambalare.				Ambalare pentru transport.				
	В	С	Α	ı	d's	B1	C1	A1	ili
621403	197	47	167	385	1	-	-	-	-